

חשוב! התקנה של יותר ממכשיר אחד, דורש אישור הנדסי.

No	Instrument Description	Model	Power Spec	Vendor	Back-up Batt.
1	Battery Charger Monitor, Defibrillator	PD4420	220 VAC 100W Battery only	ZOLL	Y
2	Bpap	Isleep 22	220VAC/50Hz	BREAS MEDICAL	N
3	Bpap	Isleep 25	220VAC/50Hz	BREAS MEDICAL	N
4	Bpap	Vivo 30	100 to 240VAC 50/60 Hz	BREAS MEDICAL	N
5	Bpap	Vivo 40	100 to 240VAC 50/60 Hz	BREAS MEDICAL	N
6	Bpap	P/N 332108	220VAC 50Hz, 400mW	RESPIRONICS	Y
7	Breather	OXYLOG 2000 P/N 8413950	220VAC, 50Hz 12-30VDC	DRAGER MEDIZINTECHNIK	Y
8	Charger	UL60601	(20W, 0.12A)		N
9	Cpap	Isleep 10	220VAC/50Hz	BREAS MEDICAL	N
10	Cpap	Isleep 201	220VAC/50Hz	BREAS MEDICAL	N
11	Cpap	PV102	100 to 240VAC 50/60 Hz	BREAS MEDICAL	N
12	Cpap	Hospital	220 VAC 50 Hz	ELECTRICAL NEBULIZER	N
13	Cpap	KXS	100 to 240VAC 50/60 Hz	KAERYYS	N
14	Cpap	Goodknight 420,	100 to 240VAC 50/60 Hz	PURITAN BENNETT	Y
15	Cpap	Goodknight 418,	100 to 240VAC 50/60 Hz	PURITAN BENNETT	Y
16	Cpap	S8 Elite Stile	220 VAC 50 Hz	RESMED	N
17	Defibrillator		220 VAC 50 Hz	GUIDANT	Y
18	Defibrillator	Heartstart MRX	220 VAC/50 Hz	PHILIPS	Y
19	Feeding cump	Kangaroo 330	220V	SHERWOOD MEDICAL	Y
20	Feeding pump	Kangaroo 2100	220VAC/50Hz 0.5MAX	SHERWOOD MEDICAL	N
21	Incubator	LEC 960	Battery Charger Voltage: 100 to 240VAC. Not allowed	LEC INSTRUMENTS	N
22	Infuccheck Volume Pump	1501EE	220V	IVAC	Y
23	Infusion pump	IMED 960	110V	IMED corporation	Y
24	Infusion Pump	Bodytwins 121	100 to 240VAC 50/60Hz	C.M.G LTD	Battery Li-ion 7.2V, 1450MA (7.5 hours-
25	Inhalation	Every - Aerosol A Micropistone	Adaptor 15VDC, (plus-inside) 1.3A max	INDOLOR AIRTHERAPY	ARTSANA 120VDC 800MA
26	Inverter	DC to AC Power	Input: 12VDC Output: 120VAC	APC	N
27	Life Compressor	Nebulizer	220VAC/50Hz	VEGA-TEC TAIWAN	N
28	Life Style Battery Charger	GTM21097-5024 P/N TR9C12 100CCP- Y-MED	24VAC/2A, 100-240VAC 50/60Hz		Y
29	Medsystem III	I VAC	110/220 VAC, 60Hz	ALARIS	Y
30	Monitor	310	220V, 0.4 Amps	CARDIO PAK 2000, (Mennen Medical)	Y
31	Monitor/Defibrillator	CP 200	220V	CARDIO PAK	Y
32	Monitor/Defibrillator	M3535A Hartstart	100-240VAC 50-60/410 1-0, 46A 18VDC 5A, 90W Batt. 14.8v 6.0AH	PHILIPS	Y
33	Monitor/Defibrillator	Lifepack 10	No External	PHYSIO CONTROL	Battery only
34	Oxyen	ECLIPSE (1000)	100 to 240VAC 50/60 Hz	SEQUAL	(1.5-3hrs - continuous) (2-4.3hrs -
35	Pulse Oximeter	N-100E	220V, 0.3 Amps	NECLOR	Y
36	Pulse Oximeter	3740	220V	OHMEDA	Y

37	Respirator	Sullivan	110-240VAC (200VA)	RESMED	N
38	Respirator INVENT	201 IC	110-240VAC, 50/60 Hz, 1.6 Amp	VERSAMED	Y
39	Suction Pump	753790	120VAC/.75A, 60Hz 240VAC/.5A , 50Hz	AERAS	Y
40	Suction pump	7304D	220V	DEVILBISS	Y
41	Suction Pump	Vacuaid 7304F	220VAC	DEVILBISS	N
42	Suction Unit	91-8850	115V	PORT A VAC	Y
43	Syringe Pump	8714843	220V, 50/60H	GRASEBY	Y
44	Vacuum Pump	ACCUVAC Rescue	12-13.8 VDC	WEINMANN	Y
45	Ventilator	LTV 900	220 VAC 50 Hz	VIASYS HEALTHCARE	N
46	Ventilator	HT50H1	100-240VAC max.2A 50/60/400 Hz 12-30VDC max 12A		Y
47	Volume ventilation	ELISEE 150	220 VAC 50 Hz	SAIM MEDICAL or RESMED	N
48	volumetric infusion pump	927	220V, 100MA	IMED corporation	Y
49	Volumetric Infusion Pump	PC-1	220 VAC 50/60 Hz0.2A x 2	IMED Corporation	Y
50		VOYAGE	220 VAC 50 Hz	MARKOS MEFSV SPA (υαα)	N
51		MICRO AIR	No external power for recharge	OMERON MATSUSAKA	2x1.5VDC (3VDC,1.5W)